



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 061585 0036 Rev. 00

Manufacturer:

B. Braun Medical AG

Seesatz 17
6204 Sempach
SWITZERLAND

**Authorized
Representative:**

B. Braun Melsungen AG
Carl-Braun-Str. 1, 34212 Melsungen, GERMANY

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result. The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis. The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with.
For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G10 061585 0036 Rev. 00

Report No.: 713183172

Valid from: 2021-03-15

Valid until: 2026-03-14

Christoph Dicks
Head of Certification/Notified Body

Issue date: 2021-03-15



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 061585 0036 Rev. 00

Classification: IIa
Device Group: D99 - DISINFECTANTS AND ANTISEPTICS FOR MEDICAL DEVICES - OTHERS
Intended Purpose: Surface disinfection of non-invasive medical devices

Classification: IIa
Device Group: D0701 - DISINFECTANTS, MEDICAL DEVICES, ETHANOL
Intended Purpose: Surface disinfection of non-invasive medical devices

Classification: IIa
Device Group: D02010102 - DISINFECTANTS, MEDICAL DEVICES, CHLORHEXIDINE, HYDROALCOHOLIC SOLUTION
Intended Purpose: Surface disinfection of non-invasive medical devices

Classification: IIb
Device Group: D01010102 - DISINFECTANTS, MEDICAL DEVICES, GLUTARALDEHYDE, ACID SOLUTION
Intended Purpose: Instrument disinfection of invasive medical devices

The validity of this certificate depends on conditions and/or is limited to the following: -