

[ H.-P. WERNER | BORNHÖVEDSTRASSE 78 | 19055 SCHWERIN ]

Fresenius Medical Care Deutschland GmbH  
Else-Kröner-Straße 1

61352 Bad Homburg

c/o

HygCen Germany GmbH  
Bornhövedstrasse 78  
19055 Schwerin

Phone: +49 (0) 385 5682 65

Fax: +49 (0) 385 5682 67

Email: [info@hygcen.de](mailto:info@hygcen.de)

Web: [www.hygcen.de](http://www.hygcen.de)

2018-01-23

**Puristeril 340**

**EN 14348**

**Mycobactericidal (*Mycobacterium chimaera*) activity**

## J U D G E M E N T

After testing the disinfectant **Puristeril 340** according to EN 14348 I give the following statement of the results out of the test report dated 2015-02-03 (SN 17700 II):

According to EN 14348, the batch VER 05200 of the product **Puristeril 340** possesses a mycobactericidal activity under clean conditions (0.3g/l bovine albumine) at 20°C in 5 minutes when diluted at 3.0% (v/v) in water of standardised hardness against the test strain *Mycobacterium chimaera* DSM 44623.

Prof. Dr. med. H.-P. Werner