

Dr. Ralph DerraÖffentlich bestellter und vereidigter Sachverständiger für
Verpackungsmaterialien, Boden- und Luftanalysen;
Sachverständiger in der Wasseranalytik

9 January 2018

Dr. Dr/Za-mue

TEST CERTIFICATE**936 P 18**

for Messrs: Ahlstrom-Munksjö Specialties
ZI de l'Abbaye - 364, Impasse Louis Champin
38780 Pont-Évêque / FRANCE

Product: Crepe paper

Designation: Reliance® 310

The product manufactured by the above-mentioned company is intended to be used as packaging material for terminally sterilized medical devices which are to be sterilized by steam (134 °C), by ethylene oxide, by gamma irradiation (50 kGy) or by formaldehyde.

The sample material submitted was tested according to

DIN EN ISO 11607-1: 2017 section 5.2.3, analysis according to DIN 58 953-6:2016,
DIN EN 868-2: 2017 section 4.2.1 as well as section 4.2.2.2.

The results are summarized in the test report no. 2172/29-1 of 17 November 2017 issued by ISEGA Forschungs- und Untersuchungsgesellschaft mbH Aschaffenburg, Germany.

The product "Reliance® 310" meets the requirements for packaging for terminally sterilized medical devices according to

DIN EN ISO 11607-1: 2017 section 5.1.6 a) microbial barrier,
section 5.1.6 c) physical and chemical properties,
section 5.1.6 e) compatibility with respect to the intended
sterilization processes

with regard to the above-mentioned sterilization processes and specification.

**(Zahn)**Head of
Physical Material Testing