

Pont-Audemer, July 1st, 2018

RELIANCE® Sterile Barrier System
Final Pack Test at TNO, The Netherlands

To whom may concerns,

Ahlstrom-Munksjö Specialties Pont-Audemer certifies that the RELIANCE® Sterile Barrier System range has been tested regarding the final pack test and showed a results of > 99.99 % retention.

This test has been developed by RIVM, Dutch institute for public health and the environment. The testing has been performed at TNO, Research and Knowledge centre in The Netherlands.

Reports are available upon request

Kind Regards,



Baptiste Trocquet

Quality Manager

Ahlstrom-Munksjö S.p.A. France
15, RUE DES PAPETIERS
27500 PONT-AUDEMER
CAPITAL SOCIAL 12 683 712 €
SIRET 352 703 201 0001
TVA / VAT FR533527032
EORI FR3527032010000

Contact office:

Ahlstrom-Munksjö Specialties

Usine de Pont-Audemer
15, Rue des Papetiers
F-27501 Pont-Audemer, France

T: +33 (0) 2 3241 6100
F: +33 (0) 2 3241 4431