



Product Service

CERTIFICATE

No. B 12 06 13615 111

Holder of Certificate: Semperit Technische Produkte GmbH

Modcenterstraße 22

1031 Wien

AUSTRIA

Production Facility(ies):

12175

Certification Mark:



Product: Medical Gloves for single use

Model(s): Latex Powder Free Surgical Gloves
"Sempermed Supreme"

Parameters:

Size

Art.-No.

5.5

8227 51 521, 8227 51 525, 8227 51 529

6

8227 51 601, 8227 51 605, 8227 51 609

6.

8227 51 621, 8227 51 625, 8227 51 629

7

8227 51 701, 8227 51 705, 8227 51 709

7.

8227 51 721, 8227 51 725, 8227 51 729

8

8227 51 801, 8227 51 805, 8227 51 809

8,5

8227 51 821, 8227 51 825, 8227 51 829

9

8227 51 901, 8227 51 905, 8227 51 909

Tested according to: EN 455-1:2000
EN 455-2/A1:2011
EN 455-3:2006
EN 455-4:2009

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: THA0169211T074-7

AA A₂

Date, 2012-06-12

(Peter Arndt Porzler)

