

Declaration of Conformity

for the **gke Steri-Record**[®] - Doku-Labeler

The **gke Steri-Record**[®] - Doku-Labelers are available in the following versions:

Art. No.	Description	Version	Application
240-840	Doku-Labeler-A*	alphanumeric	for all gke Steri-Record [®]
240-850	Doku-Labeler-N**	numeric	3-line documentation labels

* the whole first line can print alphanumeric digits

** the first three digits of the first line can print alphanumeric digits, the remaining ones print numbers

The hand labeler device containing three printing lines with 12 digits each is manufactured of various plastic and stainless steel materials where self-adhesive labels and ink roll and a printing head with rubber letters are implemented.

The hand labeler is exclusive part of **gke-Steri-Record**[®] documentation system for correct application it should be used only in combination with original **gke-Steri-Record**[®] double self-adhesive documentation labels art. no. 240-871 to 240-874, 242-875, 242-885, 243-873, 243-883, 244-871, 244-881, 240-861 to 240-864, 000-863, 000-873, 000-875.

All hand labeler devices are no medical devices and do not require a CE-mark.

This document certifies that the above performance criteria and the **gke** test requirements for quality control are met. The continuous quality is guaranteed by our quality management system according to EN ISO 13485:2003**.



Waldems-Esch, 2010-10-06

gke-GmbH
Auf der Lind 10
D-65529 Waldems-Esch
Germany

Dipl.-Ing. Dr. Ulrich Kaiser
R & D-Manager

753-009-EN-PL V03