



Issued to:

Abena A/S
Egelund 35
6200 Aabenraa
Denmark

Notified Body: 2777

SATRA customer number: P1144

EU Type-Examination Certificate

Certificate number: 2777/11578-03/E01-01

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation. It has been issued Under Module B of Regulation 2016/425 on personal protective equipment. This product group has been shown to satisfy the applicable essential health and safety requirements as a Category III product.

Product reference:

Abena Classic Protect
nitrile gloves powder free

Description:

≥ 2.2 mil Powder Free Nitrile disposable five fingered glove
Available in a standard minimum 240mm length or a longer cuff variant of 280mm,
available as non-sterile.

Sizes:

- 6 (XS)
- 7 (S)
- 8 (M)
- 9 (L)
- 10 (XL)

Classification:

EN ISO 374-1:2016 + A1: 2018/ Type B	Level	EN ISO 374-4:2019 Degradation (%)
40% Sodium Hydroxide (K)	6	-25.7
30% Hydrogen Peroxide (P)	2	44.8
37% Formaldehyde (T)	5	-17.1

EN ISO 374-5:2016

	Level
Protection against bacteria and fungi	Pass
Protection against viruses	Pass

Standards/Technical specifications applied:

EN ISO 21420:2020; EN ISO 374-1:2016+A1:2018; EN ISO 374-5:2016; EN 421: 2010 (excluding clause 4.3)

Technical reports/Approval documents:

SATRA: CHM0264532/1746/EN/A; CHM0264532/1746/EN/B; CHM0272837/1827/LC/JH/A/Issue 2; SPC0232012/1505 Issue 2; CHM0268337/1810/EN; CHM0259706/1728/SPT; CHM0260437/1730/EN; CHM0264532/1746/EN/C; SPC0339493/2242/1; CHM0334576/2229/CC; CHM0334576/2229/SPT Issue 2; SPC0337444/2236/2; CHM0350835/2323/2; CHM0350835/2323/1
SGS: CH:TX:1342045578; CH:TX:1342045581; CH:TX:1342045582; CH:TX:1342045583; CH:TX:1342045584; CH:TX:1342045585

Signed on behalf of SATRA:

Geoff Graham

Date of issue: 16/10/2023
Expiry date: 01/09/2028