



Issued to:

Abena A/S
Egelund 35
6200 Aabenraa
Denmark

Notified Body: 2777

SATRA customer number: P1720

EU Type-Examination Certificate

Certificate number: 2777/17447-02/E01-01

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:

Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

Product reference:

Description:

1999902086-1999902090(XS-XL)
1999905271-1999905275(XS-XL)

Nitrile Medical Examination Gloves
Colour: Blue

Sizes:

5-6/XS
6-7/S
7-8/M
8-9/L
9-10/XL

Classification:

EN ISO 374-1:2016+A1:2018/Type B

Level

EN ISO 374-4:2019 Degradation %

(K) Sodium hydroxide 40%
(P) Hydrogen peroxide 30%
(T) Formaldehyde 37%

6
2
3

-11.5
-9.5
7.4

EN ISO 374-5:2016

Level

Protection against Bacterial and Fungi
Protection against Viruses

Pass
Pass

Standards/Technical specifications applied:

EN ISO 21420:2020; EN ISO 374-1:2016+A1:2018; EN ISO 374-5:2016

Technical reports/Approval documents:

SATRA: CHT0291374/1944, CHM0291937/1946/JH, CHT0301241/2033
SGS: CH:TX:9420026599-1, CH:TX:9420026316-1, CH:TX:9420020333, CH:TX:9420029243
CTC: S200908976_2
TUV: 721655656

Signed on behalf of SATRA:

Quincey Brown

Date of issue: 07/12/2021
Expiry date: 15/07/2026